

America

CERTIFICATE

U8 066044 0032 Rev. 05

Holder of Certificate: **Cognex Corporation**
One Vision Drive
Natick MA 01760
USA

Certification Mark:




Product: **Laboratory Equipment**

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL for USA and a Standards Council of Canada ISO/IEC 17065 accredited Certification body for Canada.

Test report no.: 713311098

Date, 2024-03-14

(Ralph Fischer)



America

CERTIFICATE

U8 066044 0032 Rev. 05

Model(s): Regulatory Model 50201
Regulatory Model 50203 (IS3D-L4xxx, where “xxx” may be 033, 050, 100 or 300 representing divergence angle)

Brand Name: COGNEX

Tested according to: UL 61010-1:2012/R:2023-06
CSA C22.2 No. 61010-1:2012/U3:2023-06
CSA E60825-1:2015

Parameters:

Regulatory Model 50203 (IS3D-L4xxx, where “xxx” may be 033, 050, 100 or 300 representing divergence angle)

Rated Input voltage:	24VDC
Rated Input Current:	750 mA
Protection Class:	III
Laser Class:	2M

Regulatory Model 50201

Rated Input voltage:	24VDC
Rated Input Current:	750 mA
Protection Class:	III
Laser Class:	2